

PROGRESSA[®] SMART BED SYSTEM

Empower patients and caregivers with in-bed therapies to help get your most critical patients better sooner.



STANDARD FEATURES TO HELP OPTIMIZE PATIENT SAFETY AND IMPROVE CAREGIVER SATISFACTION

- SlideGuard™ Frame Technology
- Patient Controls Backlighting
- FlexAfoot™ Retractable Foot Control
- Advanced MicroClimate® Technology
- 30°/45° Head of Bed Alarm
- In-Bed Scale
- Line Managers
- Three Mode Bed Exit Alarm
- Caregiver Pendant
- Caregiver Foot Controls
- Night Light
- Integrated Oxygen Tank Holder
- Accessory Outlet
- Hands Free Emergency CPR & Trendelenburg
- Point of Care™ Siderails Control
- Both Siderails with Point-of-Care Touch Screen
- Obstacle Detect™ System at Both Sides of the Bed
- Drainage Bag Holders
- Four IV Sockets

OPTIONAL FEATURES TO SUPPORT THERAPY-FOCUSED TECHNOLOGIES

- Prevention Surface
- Therapy Surface
- Pulmonary Surface
- Percussion & Vibration Therapy (P&V)
- Bed Exit Alarm to Nurse Call
- 5th Wheel Steering
- IntelliDrive® Powered Transport Steering
- Transport Shelf
- Wireless Connectivity
- Permanent IV Pole
- Full Chair Egress
- StayInPlace™ Patient Migration Management System

TECHNICAL SPECIFICATIONS

Total Length			
Foot Section Extended	98" (248.9 cm)	Safe Working Load	650 lb (295 kg)
Foot Section Retracted	88" (223.5 cm)	Maximum Patient Weight	500 lb (227 kg)
Maximum Width		C-Arm Clearance	7 - 15" (17.8-38.1 cm)
Side Rails Down	40.5" (103 cm)	Cleanability/Infection Control	IPX4
Side Rails Up	40.5" (103 cm)	Radiolucent Sleep Deck	Standard 17.7" x 23" (45 cm x 58.5 cm)
Maximum Side Rail Height Above Sleep Deck	17" (43.2 cm)	Voltages	100/110/115/120/127/220/230/240
Highest Position Top of Seat Section to Floor	35.7" (90.7 cm)	Mattress	
Lowest Position Top of Seat Section to Floor	16.5" (41.9 cm)	Prevention Surface	
Minimum Underbed Clearance	4.3" (10.9 cm) w/o IntelliDrive® Powered Transport	Mattress Width	35" (88.9 cm)
Standard Caster Size	6" (15.2 cm)	Mattress Length	84" (213 cm)
Articulations		Mattress Height	7.1" (18.1 cm)
Head Section	0° - 67° 77° with Chair Egress	Mattress Weight	31 lb (14.1 kg)
Thigh Section	0° - 30°	Therapy Surface	
Foot Section	0° - 45°	Mattress Width	35.5" (90.2 cm)
Trendelenburg Position (maximum)	75° with Chair Egress	Mattress Length	84" (213 cm)
Reverse Trendelenburg Position (maximum)	13° 18°	Mattress Height	8" (20.32 cm)
		Mattress Weight	45 lb (20.5 kg)
		Pulmonary Surface	
		Mattress Width	35.5" (90.2 cm)
		Mattress Length	84" (213 cm)
		Mattress Height	8" (20.32 cm)
		Mattress Weight	48 lb (21.8 kg)

SURFACE & FRAME CONFIGURATION OPTIONS

PROGRESSA® PREVENTION	PROGRESSA® THERAPY	PROGRESSA® PULMONARY
Prevention Surface	Therapy Surface	Choose one: - Pulmonary Surface (CLRT only) - Pulmonary Surface (CLRT & P&V)
Choose one: - Dining Chair® Position - FullChair® Position with Chair Egress - FullChair® Position with Chair Egress and StayInPlace™ Technology	Choose one: - Dining Chair® Position - FullChair® Position with Chair Egress - FullChair® Position with Chair Egress and StayInPlace™ Technology	Choose one: - Dining Chair® Position - FullChair® Position with Chair Egress - FullChair® Position with Chair Egress and StayInPlace™ Technology



hillrom.com

Hill-Rom reserves the right to make changes without notice in design, specifications and models. The only warranty Hill-Rom makes is the express written warranty. The Progressa Bed may be used in a variety of settings including, but not limited to, acute care, including critical care, step down/progressive care, medical/surgical, high acuity sub-acute care, post anaesthesia care unit (PACU), and sections of the emergency department (ED). Medical device (93/42/EEC): Class Im, Class IIa for therapy and pulmonary surfaces. Notified Body: Polskie Centrum Badan I Certyfikacji S.A.(PCBC) CE1434. Scale Classification (EN 45501): Class III. Manufacturer: Hill-Rom Holdings, Inc. - 1069 State Route 46 East - Batesville, IN 47006 - USA. European Representative: Hill-Rom S.A.S. - BP 14 ZI du Talhouët 56330 Pluvigner France. This medical device is a regulated health product which, pursuant to such regulation bears a CE mark. Hill-Rom recommends that you carefully read the detailed instructions for safe and proper use included in the documents accompanying the medical devices. The personnel of healthcare establishments are responsible for the proper use and maintenance of these medical devices.

Hill-Rom reserves the right to make changes without notice in design, specifications and models. The only warranty Hill-Rom makes is the express written warranty extended on the sale or rental of its products.